



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
-----------------	-------------	----------------------	---------------------	------------------

10/567,495

07/31/2006

James R. Hewit

CAF-35102/03

1070

25006

7590

03/17/2008

GIFFORD, KRASS, SPRINKLE, ANDERSON & CITKOWSKI, P.C

PO BOX 7021

TROY, MI 48007-7021

EXAMINER

STOUT, MICHAEL C

ART UNIT

PAPER NUMBER

3736

MAIL DATE

DELIVERY MODE

03/17/2008

PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Office Action Summary	Application No. 10/567,495	Applicant(s) HEWIT ET AL.	
	Examiner MICHAEL C. STOUT	Art Unit 3736	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 1 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☐ Responsive to communication(s) filed on ____.
- 2a) ☐ This action is **FINAL**. 2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-44 is/are pending in the application.
- 4a) Of the above claim(s) ____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) ____ is/are allowed.
- 6) ☐ Claim(s) ____ is/are rejected.
- 7) ☐ Claim(s) ____ is/are objected to.
- 8) ☒ Claim(s) 1-44 are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on ____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. ____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- | | |
|--|---|
| 1) <input type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input type="checkbox"/> Interview Summary (PTO-413) |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | Paper No(s)/Mail Date. ____. |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08) | 5) <input type="checkbox"/> Notice of Informal Patent Application |
| Paper No(s)/Mail Date ____. | 6) <input type="checkbox"/> Other: ____. |

DETAILED ACTION

Election/Restrictions

1. Restriction is required under 35 U.S.C. 121 and 372.

This application contains the following inventions or groups of inventions which are not so linked as to form a single general inventive concept under PCT Rule 13.1.

In accordance with 37 CFR 1.499, applicant is required, in reply to this action, to elect a single invention to which the claims must be restricted.

Group 1, claim(s) 1-41, drawn to a palpation device.

Group 2, claim(s) 42, drawn to a palpation simulation device.

Group 3, claim(s) 43, drawn to a method of palpating a body part.

Group 4, claim(s) 44, drawn to a method of simulating palpation of a body part.

2. The inventions listed as Groups 1-4 do not relate to a single general inventive concept under PCT Rule 13.1 because, under PCT Rule 13.2, they lack the same or corresponding special technical features for the following reasons:

DI: US-A-5 769 791 (GOLDBERGER DANIEL S ET AL) 23 June 1998 (1998-06-23)

D2:US-B-6 345 1941 (ZACH REUVEN D ET AL) 5 February 2002 (2002-02-05)

Document D1 (the references in parentheses applying to this document) relates to a palpation device comprising

a first palpation assembly including a palpation member (grasper member 201, col. 18, l. 12 and Fig. 6).and a light source (fiber 245, col. 18, l. 20 and Fig. 6); and

a second palpation assembly including a palpation member (grasper member 203, col. 18, l. 12 and Fig. 6) and light detecting means for detecting light emitted by the light source (fiber 246, col. 18, l. 20 and Fig. 6) and generating an image of a body part disposed between the first and second palpation members (col. 5, l. 42-43);

wherein at least one of the first and second palpation members is movable with respect to the other member, to palpate a body part disposed there between (col. 18, l. 9-21 and Fig. 6).

Since the device of D1 is explicitly suitable to generate images (D1, col. 5, l. 42-43), all structural features mentioned in or derivable from present claim 1 are known from D1. Furthermore, the applicant is requested to note, that also the breast imaging device of D2 (e.g. col. 6, l. 45-54 and claims 1 and 4) contains all features of present claim 1 and is therefore considered to be at least suitable for the claimed purpose. Present claim 1 thus lacks novelty so that the requirement of Art. 33(2) PCT is not met.

The same reasoning applies, *mutatis mutandis*, to the subject-matter of the corresponding independent claim 42, which therefore is also considered not new in the sense of Art. 33(2) PCT.

Regarding claims 43 and 44 D1 discloses a device which simulates the palpation of tissue comprising the limitation of the claims 1 and 42 further comprising the generation of an image of the body part which is taught by D1 (see Column 5, Lines 40-50).

There is a clear lack of unity of the invention because the common matter of the independent claims is well known and the remaining subject matter of each claim differs from that of the others without there being any unifying novel inventive concept.

3. Applicant is reminded that upon the cancellation of claims to a non-elected invention, the inventorship must be amended in compliance with 37 CFR 1.48(b) if one or more of the currently named inventors is no longer an inventor of at least one claim remaining in the application. Any amendment of inventorship must be accompanied by a request under 37 CFR 1.48(b) and by the fee required under 37 CFR 1.17(i).

4. The examiner has required restriction between product and process claims. Where applicant elects claims directed to the product, and the product claims are subsequently found allowable, withdrawn process claims that depend from or otherwise require all the limitations of the allowable product claim will be considered for rejoinder. All claims directed to a nonelected process invention must require all the limitations of an allowable product claim for that process invention to be rejoined.

In the event of rejoinder, the requirement for restriction between the product claims and the rejoined process claims will be withdrawn, and the rejoined process

Art Unit: 3736

claims will be fully examined for patentability in accordance with 37 CFR 1.104. Thus, to be allowable, the rejoined claims must meet all criteria for patentability including the requirements of 35 U.S.C. 101, 102, 103 and 112. Until all claims to the elected product are found allowable, an otherwise proper restriction requirement between product claims and process claims may be maintained. Withdrawn process claims that are not commensurate in scope with an allowable product claim will not be rejoined. See MPEP § 821.04(b). Additionally, in order to retain the right to rejoinder in accordance with the above policy, applicant is advised that the process claims should be amended during prosecution to require the limitations of the product claims. **Failure to do so may result in a loss of the right to rejoinder.** Further, note that the prohibition against double patenting rejections of 35 U.S.C. 121 does not apply where the restriction requirement is withdrawn by the examiner before the patent issues. See MPEP § 804.01.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to MICHAEL C. STOUT whose telephone number is (571)270-5045. The examiner can normally be reached on M-F 7:30-5:00 Alternate (Fridays).

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Max Hindenburg can be reached on 571-272-4726. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Art Unit: 3736

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

/M. C. S./
Examiner, Art Unit 3736

/Max Hindenburg/
Supervisory Patent Examiner, Art Unit 3736